California Medical Device Recall Information

Recall Name
Bausch and Lomb Recalls 27G Sterile Cannula Packed in Bausch and Lomb Amvisc 1.2% Sodium Hyaluronate and Amvisc Plus 1.6% Sodium Hyaluronate Ophthalmic Viscosurgical Devices Due to the Possibility of the Cannula Leaking or Detaching

Recall Date	Product Description	Recalling Firm	Recall Reason
11/14/12	27G Sterile Cannula Component packed in two **Bausch and Lomb** products (see below)	**Bausch and Lomb, Inc.** Rochester, NY	*Cannula suspected of leaking or detaching from the syringe*
Recall Class	**Product Identification**	**Distribution**	**Affected Dates**
I	Bausch and Lomb 27G Sterile Cannulas are found in: • Amvisc 1.2% Sodium Hyaluronate **(Models: 59051, 59081, 59051L, 59081L)** • Amvisc Plus 1.6% Sodium Hyaluronate Ophthalmic Visco-surgical device **(Models: 60081, 60051, 60051L, 60081L)** List of Suspected Lots Recalled	**CA**, nationwide	Manufactured between December 2009 and August 2012

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm336413.htm